

第1表 アルコール分温度補正表(1)

測定温度	測定度数									
	1	2	3	4	5	6	7	8	9	10
0	1.2	2.3	3.3	4.3	5.4	6.5	7.6	8.7	9.8	11.0
1	1.3	2.3	3.3	4.4	5.4	6.5	7.6	8.7	9.8	11.0
2	1.3	2.4	3.4	4.4	5.5	6.6	7.6	8.7	9.9	11.0
3	1.4	2.4	3.4	4.5	5.5	6.6	7.6	8.7	9.9	11.0
4	1.4	2.4	3.5	4.5	5.5	6.6	7.7	8.7	9.9	11.0
5	1.4	2.4	3.5	4.5	5.5	6.6	7.7	8.7	9.8	10.9
6	1.4	2.4	3.5	4.5	5.5	6.6	7.6	8.7	9.8	10.9
7	1.4	2.4	3.4	4.5	5.5	6.6	7.6	8.7	9.8	10.8
8	1.4	2.4	3.4	4.4	5.5	6.5	7.6	8.6	9.7	10.8
9	1.3	2.4	3.4	4.4	5.4	6.5	7.5	8.6	9.6	10.7
10	1.3	2.3	3.3	4.4	5.4	6.4	7.5	8.5	9.5	10.6
11	1.3	2.3	3.3	4.3	5.3	6.3	7.4	8.4	9.5	10.5
12	1.2	2.2	3.2	4.2	5.3	6.3	7.3	8.3	9.4	10.4
13	1.1	2.2	3.2	4.2	5.2	6.2	7.2	8.2	9.2	10.3
14	1.1	2.1	3.1	4.1	5.1	6.1	7.1	8.1	9.1	10.1
15	1.0	2.0	3.0	4.0	5.0	6.0	7.0	8.0	9.0	10.0
16	0.9	1.9	2.9	3.9	4.9	5.9	6.9	7.9	8.9	9.9
17	0.8	1.8	2.8	3.8	4.8	5.8	6.8	7.7	8.7	9.7
18	0.7	1.7	2.7	3.7	4.7	5.6	6.6	7.6	8.6	9.6
19	0.6	1.6	2.6	3.6	4.5	5.5	6.5	7.5	8.4	9.4
20	0.5	1.5	2.5	3.4	4.4	5.4	6.3	7.3	8.3	9.2
21	0.4	1.3	2.3	3.3	4.3	5.2	6.2	7.1	8.1	9.0
22	0.2	1.2	2.2	3.2	4.1	5.1	6.0	7.0	7.9	8.9
23	0.1	1.1	2.0	3.0	4.0	4.9	5.9	6.8	7.7	8.7
24		0.9	1.9	2.9	3.8	4.8	5.7	6.6	7.6	8.5
25		0.8	1.7	2.7	3.6	4.6	5.5	6.4	7.4	8.3
26		0.6	1.6	2.5	3.5	4.4	5.3	6.3	7.2	8.1
27		0.5	1.4	2.4	3.3	4.2	5.2	6.1	7.0	7.9
28		0.3	1.2	2.2	3.1	4.0	5.0	5.9	6.8	7.7
29		0.1	1.1	2.0	2.9	3.8	4.8	5.7	6.6	7.4
30			0.9	1.8	2.7	3.7	4.6	5.5	6.3	7.2
31			0.7	1.6	2.5	3.4	4.3	5.2	6.1	7.0
32			0.5	1.4	2.3	3.2	4.1	5.0	5.9	6.8
33			0.3	1.2	2.1	3.0	3.9	4.8	5.7	6.5
34			0.1	1.0	1.9	2.8	3.7	4.6	5.4	6.3
35				0.8	1.7	2.6	3.5	4.3	5.2	6.0

注) この表はJIS B 7548:2009(酒精度浮ひょう)附属書A(規定)「国際アルコール表」及び「国際法定計量機関で採択された国際アルコール表(訳編 工業技術院計量研究所1977)」のTable Iに基づき作成されたものであり、これらに基づき製造された酒精度浮ひょうに適用できる。

第1表 アルコール分温度補正表(2)

測定度	測定度									
	11	12	13	14	15	16	17	18	19	20
0	12.2	13.4	14.7	16.0	17.4	18.8	20.2	21.7	23.1	24.5
1	12.2	13.4	14.7	16.0	17.3	18.6	20.0	21.4	22.8	24.2
2	12.2	13.4	14.6	15.9	17.2	18.5	19.8	21.2	22.5	23.9
3	12.2	13.3	14.6	15.8	17.1	18.3	19.6	21.0	22.3	23.5
4	12.1	13.3	14.5	15.7	16.9	18.2	19.4	20.7	22.0	23.2
5	12.1	13.2	14.4	15.6	16.8	18.0	19.2	20.5	21.7	22.9
6	12.0	13.1	14.3	15.5	16.6	17.8	19.0	20.2	21.5	22.6
7	11.9	13.0	14.2	15.3	16.5	17.6	18.8	20.0	21.2	22.3
8	11.9	12.9	14.1	15.2	16.3	17.5	18.6	19.7	20.9	22.0
9	11.8	12.8	13.9	15.0	16.1	17.3	18.4	19.5	20.6	21.7
10	11.7	12.7	13.8	14.9	16.0	17.1	18.2	19.3	20.4	21.5
11	11.5	12.6	13.7	14.7	15.8	16.9	17.9	19.0	20.1	21.2
12	11.4	12.5	13.5	14.5	15.6	16.6	17.7	18.8	19.8	20.9
13	11.3	12.3	13.3	14.4	15.4	16.4	17.5	18.5	19.5	20.6
14	11.2	12.2	13.2	14.2	15.2	16.2	17.2	18.3	19.3	20.3
15	11.0	12.0	13.0	14.0	15.0	16.0	17.0	18.0	19.0	20.0
16	10.8	11.8	12.8	13.8	14.8	15.8	16.8	17.7	18.7	19.7
17	10.7	11.7	12.6	13.6	14.6	15.5	16.5	17.5	18.4	19.4
18	10.5	11.5	12.4	13.4	14.4	15.3	16.3	17.2	18.2	19.1
19	10.3	11.3	12.3	13.2	14.1	15.1	16.0	17.0	17.9	18.8
20	10.2	11.1	12.1	13.0	13.9	14.8	15.8	16.7	17.6	18.5
21	10.0	10.9	11.8	12.8	13.7	14.6	15.5	16.4	17.3	18.2
22	9.8	10.7	11.6	12.5	13.4	14.4	15.3	16.2	17.1	18.0
23	9.6	10.5	11.4	12.3	13.2	14.1	15.0	15.9	16.8	17.7
24	9.4	10.3	11.2	12.1	13.0	13.9	14.7	15.6	16.5	17.4
25	9.2	10.1	11.0	11.9	12.7	13.6	14.5	15.3	16.2	17.1
26	9.0	9.9	10.7	11.6	12.5	13.4	14.2	15.1	15.9	16.8
27	8.8	9.6	10.5	11.4	12.2	13.1	13.9	14.8	15.6	16.5
28	8.5	9.4	10.3	11.1	12.0	12.8	13.7	14.5	15.3	16.2
29	8.3	9.2	10.0	10.9	11.7	12.6	13.4	14.2	15.1	15.9
30	8.1	8.9	9.8	10.6	11.5	12.3	13.1	13.9	14.8	15.6
31	7.9	8.7	9.6	10.4	11.2	12.0	12.8	13.7	14.5	15.3
32	7.6	8.5	9.3	10.1	10.9	11.8	12.6	13.4	14.2	15.0
33	7.4	8.2	9.1	9.9	10.7	11.5	12.3	13.1	13.9	14.7
34	7.1	8.0	8.8	9.6	10.4	11.2	12.0	12.8	13.6	14.4
35	6.9	7.7	8.5	9.3	10.1	10.9	11.7	12.5	13.3	14.1

第1表 アルコール分温度補正表(3)

測定度	測定度									
	21	22	23	24	25	26	27	28	29	30
0	25.9	27.2	28.5	29.8	31.0	32.1	33.2	34.3	35.3	36.4
1	25.5	26.8	28.1	29.3	30.5	31.7	32.8	33.8	34.9	35.9
2	25.2	26.5	27.7	28.9	30.1	31.2	32.3	33.4	34.5	35.5
3	24.8	26.1	27.3	28.5	29.7	30.8	31.9	33.0	34.0	35.0
4	24.5	25.7	26.9	28.1	29.2	30.3	31.5	32.5	33.6	34.6
5	24.2	25.4	26.5	27.7	28.8	29.9	31.0	32.1	33.1	34.2
6	23.8	25.0	26.2	27.3	28.4	29.5	30.6	31.7	32.7	33.7
7	23.5	24.7	25.8	26.9	28.0	29.1	30.2	31.2	32.3	33.3
8	23.2	24.3	25.4	26.5	27.6	28.7	29.8	30.8	31.9	32.9
9	22.9	24.0	25.1	26.2	27.2	28.3	29.4	30.4	31.4	32.5
10	22.5	23.6	24.7	25.8	26.9	27.9	29.0	30.0	31.0	32.1
11	22.2	23.3	24.4	25.4	26.5	27.5	28.6	29.6	30.6	31.6
12	21.9	23.0	24.0	25.1	26.1	27.1	28.2	29.2	30.2	31.2
13	21.6	22.6	23.7	24.7	25.7	26.7	27.8	28.8	29.8	30.8
14	21.3	22.3	23.3	24.3	25.4	26.4	27.4	28.4	29.4	30.4
15	21.0	22.0	23.0	24.0	25.0	26.0	27.0	28.0	29.0	30.0
16	20.7	21.7	22.7	23.6	24.6	25.6	26.6	27.6	28.6	29.6
17	20.4	21.4	22.3	23.3	24.3	25.3	26.2	27.2	28.2	29.2
18	20.1	21.0	22.0	23.0	23.9	24.9	25.9	26.8	27.8	28.8
19	19.8	20.7	21.7	22.6	23.6	24.5	25.5	26.5	27.4	28.4
20	19.5	20.4	21.3	22.3	23.2	24.2	25.1	26.1	27.0	28.0
21	19.2	20.1	21.0	21.9	22.9	23.8	24.7	25.7	26.7	27.6
22	18.9	19.8	20.7	21.6	22.5	23.4	24.4	25.3	26.3	27.2
23	18.6	19.5	20.4	21.3	22.2	23.1	24.0	24.9	25.9	26.8
24	18.3	19.1	20.0	20.9	21.8	22.7	23.6	24.6	25.5	26.5
25	17.9	18.8	19.7	20.6	21.5	22.4	23.3	24.2	25.1	26.1
26	17.6	18.5	19.4	20.3	21.1	22.0	22.9	23.8	24.8	25.7
27	17.3	18.2	19.0	19.9	20.8	21.7	22.6	23.5	24.4	25.3
28	17.0	17.9	18.7	19.6	20.4	21.3	22.2	23.1	24.0	24.9
29	16.7	17.6	18.4	19.2	20.1	21.0	21.8	22.7	23.6	24.5
30	16.4	17.2	18.1	18.9	19.8	20.6	21.5	22.4	23.3	24.2
31	16.1	16.9	17.7	18.6	19.4	20.3	21.1	22.0	22.9	23.8
32	15.8	16.6	17.4	18.2	19.1	19.9	20.8	21.6	22.5	23.4
33	15.5	16.3	17.1	17.9	18.7	19.6	20.4	21.3	22.2	23.0
34	15.2	16.0	16.8	17.6	18.4	19.2	20.1	20.9	21.8	22.7
35	14.9	15.6	16.4	17.2	18.1	18.9	19.7	20.6	21.4	22.3

第1表 アルコール分温度補正表(4)

測定度	測定度									
	31	32	33	34	35	36	37	38	39	40
0	37.4	38.4	39.3	40.3	41.3	42.2	43.2	44.1	45.1	46.0
1	36.9	37.9	38.9	39.9	40.9	41.8	42.8	43.7	44.7	45.6
2	36.5	37.5	38.5	39.5	40.4	41.4	42.4	43.3	44.3	45.2
3	36.1	37.1	38.1	39.0	40.0	41.0	41.9	42.9	43.9	44.8
4	35.6	36.6	37.6	38.6	39.6	40.6	41.5	42.5	43.5	44.4
5	35.2	36.2	37.2	38.2	39.2	40.1	41.1	42.1	43.1	44.0
6	34.8	35.8	36.8	37.8	38.7	39.7	40.7	41.7	42.7	43.6
7	34.3	35.3	36.3	37.3	38.3	39.3	40.3	41.3	42.3	43.2
8	33.9	34.9	35.9	36.9	37.9	38.9	39.9	40.9	41.8	42.8
9	33.5	34.5	35.5	36.5	37.5	38.5	39.5	40.5	41.4	42.4
10	33.1	34.1	35.1	36.1	37.1	38.1	39.1	40.0	41.0	42.0
11	32.6	33.7	34.7	35.7	36.7	37.6	38.6	39.6	40.6	41.6
12	32.2	33.2	34.2	35.2	36.2	37.2	38.2	39.2	40.2	41.2
13	31.8	32.8	33.8	34.8	35.8	36.8	37.8	38.8	39.8	40.8
14	31.4	32.4	33.4	34.4	35.4	36.4	37.4	38.4	39.4	40.4
15	31.0	32.0	33.0	34.0	35.0	36.0	37.0	38.0	39.0	40.0
16	30.6	31.6	32.6	33.6	34.6	35.6	36.6	37.6	38.6	39.6
17	30.2	31.2	32.2	33.2	34.2	35.2	36.2	37.2	38.2	39.2
18	29.8	30.8	31.8	32.8	33.8	34.8	35.8	36.8	37.8	38.8
19	29.4	30.4	31.4	32.4	33.4	34.4	35.4	36.4	37.4	38.4
20	29.0	30.0	31.0	32.0	32.9	34.0	35.0	36.0	37.0	38.0
21	28.6	29.6	30.6	31.5	32.5	33.5	34.5	35.5	36.6	37.6
22	28.2	29.2	30.2	31.1	32.1	33.1	34.1	35.1	36.2	37.2
23	27.8	28.8	29.8	30.7	31.7	32.7	33.7	34.7	35.8	36.8
24	27.4	28.4	29.3	30.3	31.3	32.3	33.3	34.3	35.3	36.4
25	27.0	28.0	29.0	29.9	30.9	31.9	32.9	33.9	34.9	36.0
26	26.6	27.6	28.6	29.5	30.5	31.5	32.5	33.5	34.5	35.5
27	26.2	27.2	28.2	29.1	30.1	31.1	32.1	33.1	34.1	35.1
28	25.9	26.8	27.8	28.7	29.7	30.7	31.7	32.7	33.7	34.7
29	25.5	26.4	27.4	28.3	29.3	30.3	31.3	32.3	33.3	34.3
30	25.1	26.0	27.0	27.9	28.9	29.9	30.9	31.9	32.9	33.9
31	24.7	25.6	26.6	27.5	28.5	29.5	30.5	31.5	32.5	33.5
32	24.3	25.3	26.2	27.1	28.1	29.1	30.1	31.1	32.1	33.1
33	23.9	24.9	25.8	26.7	27.7	28.7	29.7	30.7	31.7	32.7
34	23.6	24.5	25.4	26.4	27.3	28.3	29.3	30.3	31.3	32.3
35	23.2	24.1	25.0	26.0	26.9	27.9	28.9	29.9	30.9	31.9

第1表 アルコール分温度補正表(5)

測定 温度	測定 度 数									
	41	42	43	44	45	46	47	48	49	50
0	46.9	47.9	48.8	49.8	50.7	51.6	52.6	53.5	54.5	55.4
1	46.5	47.5	48.4	49.4	50.3	51.3	52.2	53.2	54.1	55.1
2	46.2	47.1	48.1	49.0	50.0	50.9	51.9	52.8	53.8	54.7
3	45.8	46.7	47.7	48.6	49.6	50.5	51.5	52.4	53.4	54.4
4	45.4	46.3	47.3	48.3	49.2	50.2	51.1	52.1	53.1	54.0
5	45.0	45.9	46.9	47.9	48.8	49.8	50.8	51.7	52.7	53.7
6	44.6	45.6	46.5	47.5	48.5	49.4	50.4	51.4	52.3	53.3
7	44.2	45.2	46.1	47.1	48.1	49.0	50.0	51.0	52.0	52.9
8	43.8	44.8	45.7	46.7	47.7	48.7	49.6	50.6	51.6	52.6
9	43.4	44.4	45.4	46.3	47.3	48.3	49.3	50.2	51.2	52.2
10	43.0	44.0	45.0	46.0	46.9	47.9	48.9	49.9	50.9	51.8
11	42.6	43.6	44.6	45.6	46.5	47.5	48.5	49.5	50.5	51.5
12	42.2	43.2	44.2	45.2	46.2	47.1	48.1	49.1	50.1	51.1
13	41.8	42.8	43.8	44.8	45.8	46.8	47.8	48.8	49.7	50.7
14	41.4	42.4	43.4	44.4	45.4	46.4	47.4	48.4	49.4	50.4
15	41.0	42.0	43.0	44.0	45.0	46.0	47.0	48.0	49.0	50.0
16	40.6	41.6	42.6	43.6	44.6	45.6	46.6	47.6	48.6	49.6
17	40.2	41.2	42.2	43.2	44.2	45.2	46.2	47.2	48.2	49.3
18	39.8	40.8	41.8	42.8	43.8	44.8	45.8	46.9	47.9	48.9
19	39.4	40.4	41.4	42.4	43.4	44.4	45.5	46.5	47.5	48.5
20	39.0	40.0	41.0	42.0	43.1	44.1	45.1	46.1	47.1	48.1
21	38.6	39.6	40.6	41.6	42.7	43.7	44.7	45.7	46.7	47.7
22	38.2	39.2	40.2	41.2	42.3	43.3	44.3	45.3	46.3	47.4
23	37.8	38.8	39.8	40.8	41.9	42.9	43.9	44.9	46.0	47.0
24	37.4	38.4	39.4	40.4	41.5	42.5	43.5	44.5	45.6	46.6
25	37.0	38.0	39.0	40.0	41.1	42.1	43.1	44.2	45.2	46.2
26	36.6	37.6	38.6	39.6	40.7	41.7	42.7	43.8	44.8	45.8
27	36.2	37.2	38.2	39.2	40.3	41.3	42.4	43.4	44.4	45.5
28	35.8	36.8	37.8	38.8	39.9	40.9	42.0	43.0	44.0	45.1
29	35.4	36.4	37.4	38.4	39.5	40.5	41.6	42.6	43.6	44.7
30	35.0	36.0	37.0	38.1	39.1	40.1	41.2	42.2	43.3	44.3
31	34.5	35.6	36.6	37.6	38.7	39.7	40.8	41.8	42.9	43.9
32	34.1	35.2	36.2	37.2	38.3	39.3	40.4	41.4	42.5	43.5
33	33.7	34.8	35.8	36.8	37.9	38.9	40.0	41.0	42.1	43.1
34	33.3	34.4	35.4	36.4	37.5	38.5	39.6	40.6	41.7	42.7
35	32.9	34.0	35.0	36.0	37.1	38.1	39.2	40.2	41.3	42.3

第1表 アルコール分温度補正表(6)

測定度	測定度									
	51	52	53	54	55	56	57	58	59	60
0	56.4	57.3	58.3	59.3	60.2	61.2	62.2	63.1	64.1	65.0
1	56.0	57.0	58.0	58.9	59.9	60.9	61.8	62.8	63.8	64.7
2	55.7	56.7	57.6	58.6	59.5	60.5	61.5	62.5	63.4	64.4
3	55.3	56.3	57.3	58.2	59.2	60.2	61.1	62.1	63.1	64.1
4	55.0	55.9	56.9	57.9	58.9	59.8	60.8	61.8	62.8	63.7
5	54.6	55.6	56.6	57.5	58.5	59.5	60.5	61.4	62.4	63.4
6	54.3	55.2	56.2	57.2	58.2	59.1	60.1	61.1	62.1	63.1
7	53.9	54.9	55.9	56.8	57.8	58.8	59.8	60.8	61.7	62.7
8	53.6	54.5	55.5	56.5	57.5	58.5	59.4	60.4	61.4	62.4
9	53.2	54.2	55.2	56.1	57.1	58.1	59.1	60.1	61.1	62.1
10	52.8	53.8	54.8	55.8	56.8	57.8	58.7	59.7	60.7	61.7
11	52.5	53.5	54.4	55.4	56.4	57.4	58.4	59.4	60.4	61.4
12	52.1	53.1	54.1	55.1	56.1	57.1	58.1	59.0	60.0	61.0
13	51.7	52.7	53.7	54.7	55.7	56.7	57.7	58.7	59.7	60.7
14	51.4	52.4	53.4	54.4	55.4	56.4	57.4	58.4	59.3	60.3
15	51.0	52.0	53.0	54.0	55.0	56.0	57.0	58.0	59.0	60.0
16	50.6	51.6	52.6	53.6	54.6	55.6	56.6	57.7	58.6	59.7
17	50.3	51.3	52.3	53.3	54.3	55.3	56.3	57.3	58.3	59.3
18	49.9	50.9	51.9	52.9	53.9	54.9	55.9	56.9	57.9	59.0
19	49.5	50.5	51.5	52.5	53.6	54.6	55.6	56.6	57.6	58.6
20	49.1	50.2	51.2	52.2	53.2	54.2	55.2	56.2	57.2	58.2
21	48.8	49.8	50.8	51.8	52.8	53.8	54.8	55.9	56.9	57.9
22	48.4	49.4	50.4	51.4	52.5	53.5	54.5	55.5	56.5	57.5
23	48.0	49.0	50.0	51.1	52.1	53.1	54.1	55.2	56.2	57.2
24	47.6	48.6	49.7	50.7	51.7	52.7	53.8	54.8	55.8	56.8
25	47.2	48.3	49.3	50.3	51.3	52.4	53.4	54.4	55.4	56.5
26	46.9	47.9	48.9	50.0	51.0	52.0	53.0	54.1	55.1	56.1
27	46.5	47.5	48.5	49.6	50.6	51.6	52.7	53.7	54.7	55.7
28	46.1	47.1	48.2	49.2	50.2	51.3	52.3	53.3	54.4	55.4
29	45.7	46.8	47.8	48.8	49.9	50.9	51.9	53.0	54.0	55.0
30	45.3	46.4	47.4	48.4	49.5	50.5	51.6	52.6	53.6	54.7
31	44.9	46.0	47.0	48.1	49.1	50.2	51.2	52.2	53.3	54.3
32	44.6	45.6	46.7	47.7	48.7	49.8	50.8	51.9	52.9	53.9
33	44.2	45.2	46.3	47.3	48.4	49.4	50.4	51.5	52.5	53.6
34	43.8	44.8	45.9	46.9	48.0	49.0	50.1	51.1	52.2	53.2
35	43.4	44.4	45.5	46.5	47.6	48.6	49.7	50.7	51.8	52.8

第1表 アルコール分温度補正表(7)

測定度	測定度									
	61	62	63	64	65	66	67	68	69	70
0	66.0	67.0	68.0	68.9	69.9	70.8	71.8	72.8	73.7	74.7
1	65.7	66.7	67.6	68.6	69.6	70.5	71.5	72.5	73.4	74.4
2	65.4	66.3	67.3	68.3	69.2	70.2	71.2	72.2	73.1	74.1
3	65.0	66.0	67.0	68.0	68.9	69.9	70.9	71.8	72.8	73.8
4	64.7	65.7	66.7	67.6	68.6	69.6	70.6	71.5	72.5	73.5
5	64.4	65.4	66.3	67.3	68.3	69.3	70.2	71.2	72.2	73.2
6	64.0	65.0	66.0	67.0	68.0	68.9	69.9	70.9	71.9	72.9
7	63.7	64.7	65.7	66.7	67.6	68.6	69.6	70.6	71.6	72.5
8	63.4	64.4	65.3	66.3	67.3	68.3	69.3	70.3	71.2	72.2
9	63.0	64.0	65.0	66.0	67.0	68.0	69.0	69.9	70.9	71.9
10	62.7	63.7	64.7	65.7	66.7	67.6	68.6	69.6	70.6	71.6
11	62.4	63.4	64.3	65.3	66.3	67.3	68.3	69.3	70.3	71.3
12	62.0	63.0	64.0	65.0	66.0	67.0	68.0	69.0	70.0	71.0
13	61.7	62.7	63.7	64.7	65.7	66.7	67.7	68.7	69.7	70.6
14	61.3	62.3	63.3	64.3	65.3	66.3	67.3	68.3	69.3	70.3
15	61.0	62.0	63.0	64.0	65.0	66.0	67.0	68.0	69.0	70.0
16	60.7	61.7	62.7	63.7	64.7	65.7	66.7	67.7	68.7	69.7
17	60.3	61.3	62.3	63.3	64.3	65.3	66.3	67.3	68.3	69.4
18	60.0	61.0	62.0	63.0	64.0	65.0	66.0	67.0	68.0	69.0
19	59.6	60.6	61.6	62.6	63.6	64.7	65.7	66.7	67.7	68.7
20	59.3	60.3	61.3	62.3	63.3	64.3	65.3	66.3	67.4	68.4
21	58.9	59.9	60.9	62.0	63.0	64.0	65.0	66.0	67.0	68.0
22	58.6	59.6	60.6	61.6	62.6	63.6	64.7	65.7	66.7	67.7
23	58.2	59.2	60.2	61.3	62.3	63.3	64.3	65.3	66.4	67.4
24	57.8	58.9	59.9	60.9	61.9	62.9	64.0	65.0	66.0	67.0
25	57.5	58.5	59.5	60.6	61.6	62.6	63.6	64.7	65.7	66.7
26	57.1	58.2	59.2	60.2	61.2	62.3	63.3	64.3	65.3	66.4
27	56.8	57.8	58.8	59.9	60.9	61.9	62.9	64.0	65.0	66.0
28	56.4	57.4	58.5	59.5	60.5	61.6	62.6	63.6	64.6	65.7
29	56.1	57.1	58.1	59.1	60.2	61.2	62.2	63.3	64.3	65.3
30	55.7	56.7	57.8	58.8	59.8	60.9	61.9	62.9	64.0	65.0
31	55.3	56.4	57.4	58.4	59.5	60.5	61.5	62.6	63.6	64.7
32	55.0	56.0	57.0	58.1	59.1	60.2	61.2	62.2	63.3	64.3
33	54.6	55.6	56.7	57.7	58.8	59.8	60.8	61.9	62.9	64.0
34	54.2	55.3	56.3	57.4	58.4	59.4	60.5	61.5	62.6	63.6
35	53.9	54.9	56.0	57.0	58.0	59.1	60.1	61.2	62.2	63.3

第1表 アルコール分温度補正表(8)

測定度	測定度									
	71	72	73	74	75	76	77	78	79	80
0	75.7	76.6	77.6	78.6	79.5	80.5	81.4	82.4	83.3	84.3
1	75.4	76.3	77.3	78.3	79.2	80.2	81.1	82.1	83.1	84.0
2	75.1	76.0	77.0	78.0	78.9	79.9	80.9	81.8	82.8	83.7
3	74.8	75.7	76.7	77.7	78.6	79.6	80.6	81.5	82.5	83.5
4	74.5	75.4	76.4	77.4	78.3	79.3	80.3	81.2	82.2	83.2
5	74.1	75.1	76.1	77.1	78.0	79.0	80.0	81.0	81.9	82.9
6	73.8	74.8	75.8	76.8	77.7	78.7	79.7	80.7	81.6	82.6
7	73.5	74.5	75.5	76.5	77.4	78.4	79.4	80.4	81.4	82.3
8	73.2	74.2	75.2	76.2	77.1	78.1	79.1	80.1	81.1	82.0
9	72.9	73.9	74.9	75.9	76.8	77.8	78.8	79.8	80.8	81.8
10	72.6	73.6	74.6	75.6	76.5	77.5	78.5	79.5	80.5	81.5
11	72.3	73.3	74.3	75.2	76.2	77.2	78.2	79.2	80.2	81.2
12	72.0	72.9	73.9	74.9	75.9	76.9	77.9	78.9	79.9	80.9
13	71.6	72.6	73.6	74.6	75.6	76.6	77.6	78.6	79.6	80.6
14	71.3	72.3	73.3	74.3	75.3	76.3	77.3	78.3	79.3	80.3
15	71.0	72.0	73.0	74.0	75.0	76.0	77.0	78.0	79.0	80.0
16	70.7	71.7	72.7	73.7	74.7	75.7	76.7	77.7	78.7	79.7
17	70.4	71.4	72.4	73.4	74.4	75.4	76.4	77.4	78.4	79.4
18	70.0	71.0	72.0	73.1	74.1	75.1	76.1	77.1	78.1	79.1
19	69.7	70.7	71.7	72.7	73.7	74.8	75.8	76.8	77.8	78.8
20	69.4	70.4	71.4	72.4	73.4	74.4	75.5	76.5	77.5	78.5
21	69.1	70.1	71.1	72.1	73.1	74.1	75.1	76.2	77.2	78.2
22	68.7	69.7	70.8	71.8	72.8	73.8	74.8	75.8	76.9	77.9
23	68.4	69.4	70.4	71.5	72.5	73.5	74.5	75.5	76.6	77.6
24	68.1	69.1	70.1	71.1	72.1	73.2	74.2	75.2	76.2	77.3
25	67.7	68.7	69.8	70.8	71.8	72.9	73.9	74.9	75.9	77.0
26	67.4	68.4	69.4	70.5	71.5	72.5	73.6	74.6	75.6	76.6
27	67.1	68.1	69.1	70.1	71.2	72.2	73.2	74.3	75.3	76.3
28	66.7	67.7	68.8	69.8	70.8	71.9	72.9	73.9	75.0	76.0
29	66.4	67.4	68.4	69.5	70.5	71.6	72.6	73.6	74.7	75.7
30	66.0	67.1	68.1	69.1	70.2	71.2	72.3	73.3	74.3	75.4
31	65.7	66.7	67.8	68.8	69.8	70.9	71.9	73.0	74.0	75.1
32	65.3	66.4	67.4	68.5	69.5	70.6	71.6	72.7	73.7	74.7
33	65.0	66.0	67.1	68.1	69.2	70.2	71.3	72.3	73.4	74.4
34	64.7	65.7	66.7	67.8	68.8	69.9	70.9	72.0	73.1	74.1
35	64.3	65.4	66.4	67.5	68.5	69.6	70.6	71.7	72.7	73.8

第1表 アルコール分温度補正表(9)

測定度 温度	測定度									
	81	82	83	84	85	86	87	88	89	90
0	85.2	86.2	87.1	88.1	89.0	89.9	90.9	91.8	92.7	93.6
1	85.0	85.9	86.9	87.8	88.8	89.7	90.6	91.5	92.5	93.4
2	84.7	85.7	86.6	87.5	88.5	89.4	90.4	91.3	92.2	93.1
3	84.4	85.4	86.3	87.3	88.2	89.2	90.1	91.1	92.0	92.9
4	84.1	85.1	86.1	87.0	88.0	88.9	89.9	90.8	91.7	92.7
5	83.9	84.8	85.8	86.8	87.7	88.7	89.6	90.6	91.5	92.4
6	83.6	84.6	85.5	86.5	87.4	88.4	89.4	90.3	91.3	92.2
7	83.3	84.3	85.2	86.2	87.2	88.1	89.1	90.1	91.0	92.0
8	83.0	84.0	85.0	85.9	86.9	87.9	88.9	89.8	90.8	91.7
9	82.7	83.7	84.7	85.7	86.6	87.6	88.6	89.6	90.5	91.5
10	82.4	83.4	84.4	85.4	86.4	87.4	88.3	89.3	90.3	91.2
11	82.2	83.2	84.1	85.1	86.1	87.1	88.1	89.0	90.0	91.0
12	81.9	82.9	83.9	84.8	85.8	86.8	87.8	88.8	89.8	90.8
13	81.6	82.6	83.6	84.6	85.6	86.5	87.5	88.5	89.5	90.5
14	81.3	82.3	83.3	84.3	85.3	86.3	87.3	88.3	89.3	90.3
15	81.0	82.0	83.0	84.0	85.0	86.0	87.0	88.0	89.0	90.0
16	80.7	81.7	82.7	83.7	84.7	85.7	86.7	87.7	88.7	89.7
17	80.4	81.4	82.4	83.4	84.4	85.4	86.5	87.5	88.5	89.5
18	80.1	81.1	82.1	83.1	84.2	85.2	86.2	87.2	88.2	89.2
19	79.8	80.8	81.8	82.9	83.9	84.9	85.9	86.9	87.9	89.0
20	79.5	80.5	81.5	82.6	83.6	84.6	85.6	86.6	87.7	88.7
21	79.2	80.2	81.2	82.3	83.3	84.3	85.3	86.4	87.4	88.4
22	78.9	79.9	81.0	82.0	83.0	84.0	85.1	86.1	87.1	88.2
23	78.6	79.6	80.7	81.7	82.7	83.7	84.8	85.8	86.9	87.9
24	78.3	79.3	80.4	81.4	82.4	83.5	84.5	85.5	86.6	87.6
25	78.0	79.0	80.1	81.1	82.1	83.2	84.2	85.3	86.3	87.4
26	77.7	78.7	79.8	80.8	81.8	82.9	83.9	85.0	86.0	87.1
27	77.4	78.4	79.4	80.5	81.5	82.6	83.6	84.7	85.7	86.8
28	77.1	78.1	79.1	80.2	81.2	82.3	83.3	84.4	85.5	86.5
29	76.7	77.8	78.8	79.9	80.9	82.0	83.0	84.1	85.2	86.2
30	76.4	77.5	78.5	79.6	80.6	81.7	82.7	83.8	84.9	86.0
31	76.1	77.2	78.2	79.3	80.3	81.4	82.4	83.5	84.6	85.7
32	75.8	76.9	77.9	79.0	80.0	81.1	82.1	83.2	84.3	85.4
33	75.5	76.5	77.6	78.7	79.7	80.8	81.9	82.9	84.0	85.1
34	75.2	76.2	77.3	78.3	79.4	80.5	81.6	82.6	83.7	84.8
35	74.8	75.9	77.0	78.0	79.1	80.2	81.2	82.3	83.4	84.5

第1表 アルコール分温度補正表(10)

測定度	測定度									
	91	92	93	94	95	96	97	98	99	100
0	94.5	95.4	96.3	97.1	98.0	98.8	99.7			
1	94.3	95.2	96.1	96.9	97.8	98.7	99.5			
2	94.0	95.0	95.8	96.7	97.6	98.5	99.4			
3	93.8	94.7	95.6	96.5	97.4	98.3	99.2	100.0		
4	93.6	94.5	95.4	96.3	97.2	98.1	99.0	99.9		
5	93.4	94.3	95.2	96.1	97.1	98.0	98.9	99.7		
6	93.1	94.1	95.0	95.9	96.9	97.8	98.7	99.6		
7	92.9	93.9	94.8	95.7	96.7	97.6	98.5	99.4		
8	92.7	93.6	94.6	95.5	96.5	97.4	98.3	99.2		
9	92.5	93.4	94.4	95.3	96.3	97.2	98.1	99.1	100.0	
10	92.2	93.2	94.1	95.1	96.1	97.0	98.0	98.9	99.8	
11	92.0	92.9	93.9	94.9	95.8	96.8	97.8	98.7	99.7	
12	91.7	92.7	93.7	94.7	95.6	96.6	97.6	98.5	99.5	
13	91.5	92.5	93.5	94.4	95.4	96.4	97.4	98.4	99.3	
14	91.2	92.2	93.2	94.2	95.2	96.2	97.2	98.2	99.2	
15	91.0	92.0	93.0	94.0	95.0	96.0	97.0	98.0	99.0	100.0
16	90.8	91.8	92.8	93.8	94.8	95.8	96.8	97.8	98.8	99.8
17	90.5	91.5	92.5	93.5	94.6	95.6	96.6	97.6	98.6	99.7
18	90.2	91.3	92.3	93.3	94.3	95.4	96.4	97.4	98.5	99.5
19	90.0	91.0	92.0	93.1	94.1	95.2	96.2	97.2	98.3	99.3
20	89.7	90.8	91.8	92.8	93.9	94.9	96.0	97.0	98.1	99.2
21	89.5	90.5	91.6	92.6	93.7	94.7	95.8	96.8	97.9	99.0
22	89.2	90.3	91.3	92.4	93.4	94.5	95.6	96.6	97.7	98.8
23	89.0	90.0	91.1	92.1	93.2	94.3	95.3	96.4	97.5	98.6
24	88.7	89.7	90.8	91.9	93.0	94.0	95.1	96.2	97.3	98.5
25	88.4	89.5	90.6	91.6	92.7	93.8	94.9	96.0	97.1	98.3
26	88.1	89.2	90.3	91.4	92.5	93.6	94.7	95.8	96.9	98.1
27	87.9	89.0	90.0	91.1	92.2	93.3	94.5	95.6	96.7	97.9
28	87.6	88.7	89.8	90.9	92.0	93.1	94.2	95.4	96.5	97.7
29	87.3	88.4	89.5	90.6	91.7	92.9	94.0	95.2	96.3	97.5
30	87.0	88.1	89.2	90.4	91.5	92.6	93.8	94.9	96.1	97.3
31	86.8	87.9	89.0	90.1	91.2	92.4	93.5	94.7	95.9	97.1
32	86.5	87.6	88.7	89.8	91.0	92.1	93.3	94.5	95.7	96.9
33	86.2	87.3	88.4	89.6	90.7	91.9	93.1	94.3	95.5	96.7
34	85.9	87.0	88.2	89.3	90.5	91.6	92.8	94.0	95.2	96.5
35	85.6	86.8	87.9	89.0	90.2	91.4	92.6	93.8	95.0	96.3